

DBH
TLN

Department of Forensic Science
Breath Alcohol Section
Trouble Call Log
6/28/2014 To 7/7/2014

Instrument Serial Number: 010446

07-Jul-14

Date: 28-Jun-14 **Time:** 2:43 AM **Remote:** True **Tech:** DBH (21250) **Location:** Charlottesville PD

Indication: DISABLED

Problem: DISABLED. INSTRUMENT RECEIVED TWO "OUT OF TOLERANCE" MESSAGES.

Technician Response: SITE VISIT CONDUCTED.

Resolution REMOVED TO DFS-Central

BH
TL✓

**Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 7/2/2014**

Instrument Serial Number: 010446

As of 02-Jul-14

Start Date	End Date	Agency Name
30-Jun-14		DFS Central Lab
30-Apr-14	30-Jun-14	Charlottesville PD
18-Apr-14	30-Apr-14	DFS Central Lab
15-Nov-12	18-Apr-14	Meherrin River Regional Jail
19-Dec-11	15-Nov-12	DFS Central Lab
09-Dec-11	19-Dec-11	Chesterfield County PD
08-Dec-11	09-Dec-11	DFS Central Lab
25-Jul-11	08-Dec-11	Chesterfield County PD
15-Oct-10	25-Jul-11	DFS Central Lab
09-Oct-09	15-Oct-10	Chesterfield County SO
30-Apr-09	09-Oct-09	DFS Central Lab
13-Nov-08	30-Apr-09	Nelson County Sheriff's Office
19-May-08	13-Nov-08	DFS Central Lab

DBH
TLN**INTOX EC/IR II**
Quality Assurance Worksheet

Instrument Serial Number **010446** Worksheet Start Date **6/30/2014**
Location **Albemarle CO Regional Jail**
Address **160 Peregory Ln Charlottesville VA 22902**
DFS Technician **Donald Hall** License No. **21250**

☐ Laboratory☒ On-SiteSite Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) Reference Barometer (mm HG)
Reference Barometer(RB)Serial # **009109** RB Calibration Due **8/1/2014**


Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.300		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.100		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Dry gas standard Lot No. (with tank no.) ☐ Replaced dry gas standard (+O-ring)☐ Installed at Location☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument test and measurement assurance checks not run due to instrument being disabled.

Instrument Serial Number

010446

Certification Date

☐ Calibrated☐ Certified☐ Measurement Assurance Check☐ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Donald B Hall

Date

7/2/14

Issuing Analyst

Tin
msk

Date

7/2/14
7/7/14 Tm